

MEDICAL GRADE NITRILE GLOVES

PERSONAL PROTECTIVE EQUIPMENT

■ 5935 WEST 84TH ST. INDIANAPOLIS, IN 46278

■ +1 (866).500.6907

■ FAX : (800).934.2082

■ WWW.DTSUS.NET

NITRILE GLOVES

MEDICAL-GRADE, POWDER-FREE



ABOUT

Nitrile, Powder-Free gloves. Using rigorous sourcing and testing to ensure that every Nitrile glove is up to DTS standards. These examination gloves offer added value to users in terms of comfort, and enhanced protection against chemicals. These gloves have been tested for resistance to selective chemicals and chemotherapy drugs.

SPECIFICATIONS

| | |
|-----------------|-----------------|
| Material | Nitrile |
| Color | Blue |
| Inner Treatment | Chlorination |
| Outer Surface | Finger Textured |

USAGE

| | |
|--|---------------------|
| | Personal Protection |
| | Hospital / Dental |
| | Laboratory / Pharma |
| | Food / Catering |
| | Cleaning |

MEASUREMENTS

| | | |
|-------------------|-------------------------------------|---|
| Length | Median 240mm | |
| Thickness at Palm | Specification 0.05 mm 2.0 mil | Actual Production 0.08 mm 3.1 mil |
| Weight | 3.5 ± 0.2g | |
| AQL Up To | 1.5 | |

STANDARDS

| | |
|---|-------------------|
| 1 | 510(K) |
| 2 | ISO 10993-10:2010 |
| 3 | ISO 10993-5:2009 |
| 4 | ASTM D6124-06 |
| 5 | ASTM D5151-06 |
| 6 | ASTM D6319-10 |
| 7 | ASTM D6978-05 |

CERTIFICATIONS

| | |
|---|-----------------------------|
| MDD (Medical Device Directive 93/42/EEC) | Class 1 |
| PPE (Personal Protective Equipment 89/868 EEC) | Category III Complex Design |
| Food Contact (EC) No. 1935/2004 | Passed |
| Food Contact (ISO 22000) Food Safety Management | Passed |